


T-Square[®] Oral Fluid Drug Test

Oral Fluid Drug Test is a simple and convenient method of drug screening. Full observation during collection allows the testing to be performed anytime and anywhere, which greatly reduces the specimen-tampered concerns.

Because it takes time for drug to metabolize and pass through body, oral fluid is also recognized as the optimal method for drug screening.



Employment & Insurance (E&I) exempt



4432 N 109th Street • Wauwatosa • WI • 53225 414-858-8014 • www.remedynowconsulting.com





- Optional saliva saturation indicator
- Convenient rapid screening
- Test up to 12 drugs & alcohol
- E&I exempt testing panels
- Rapid results

Exempt Panels

Drug	Cutoff
Barbiturates (BAR)	20 ng/mL
Methylenedioxy-methamphetamine (MDMA)	50 ng/mL
Marijuana (THC)	40 ng/mL
Amphetamine (AMP)	50 ng/mL
Cocaine (COC)	20 ng/mL

Drug	Cutoff
Methamphetamine (MET)	50 ng/mL
Opiate (OPI)	40 ng/mL
Oxycodone (OXY)	20 ng/mL
Methadone (MTD)	30 ng/mL
Phencyclidine (PCP)	10 ng/mL

Forensic Use Panels

滋 滋 滋 滋	Alcohol (ACL)	0.02g/dL
AME AME AME AME	Amphetamine (AMP)	50ng/mL
AL AL AL	Barbiturates (BAR)	60ng/mL
	Benzodiazepines (BZO)	30ng/mL
	Buprenorphine (BUP)	5ng/mL
***	Cocaine (COC)	20ng/mL
HE HE HE HE	Cannabinoids (THC)	25ng/mL
EDDE EDDE EDDE	2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP)	20ng/mL
	Fentanyl (FTY)	30ng/mL

類 類 類 類	Ketamine (KET)	100ng/mL
雞雞雞雞	Methadone (MTD)	30ng/mL
冷計 冷計 冷計 冷計	Methamphetamine (MET/mAMP)	50ng/mL
MĚMÝ MĚMÝ MĚMÝ	Methylenedioxymethamphetamine (MDMA)	100ng/mL
XII XII XII XII	Opiate (OPI)	40ng/mL
W W W W	Oxycodone (OXY)	20ng/mL
矮矮矮矮	Phencyclidine (PCP)	10ng/mL
* * * * *	Synthetic Canabisnoids (K2)	5ng/mL

remedyNOW Consulting, LLC

4432 N 109th Street • Wauwatosa • WI • 53225 414-858-8014 • www.remedynowconsulting.com




Allow the kit and specimen to come to room temperature prior to testing.

AVOID PLACING ANYTHING IN THE MOUTH 10 MINUTES PRIOR TO TESTING.


Step 1

Remove the test cube and the sponge collector from the foil pouch by tearing at the notch. Place the test cube upright on a level surface.



Step 2


Put the sponge end of the collector on your tongue to collect oral fluid for about 7 minutes or 3 minutes if using the saturation indicator, until the color on saturation indicator strip appears RED in the indicator window.



Step 3

Open the test cube and place the fully saturated sponge collector inside the test cube. Press the sponge collector down firmly until it reaches the bottom of the test cube, then close the cube lid tightly while compressing the collector. Keep test cube upright on a flat surface and follow Step 4.

Note: Make sure the sponge collector is inserted vertically and the handle of collector is put into the clamp.



Step 4

- 1. Interpreting Alcohol Test Result:
 - Read result at 2 minutes.
 - NEGATIVE: No color changes on reaction pad.
 - POSITIVE: Compare the color of reaction pad with provided color chart to estimate the approximate blood alcohol level.
- 2. Interpreting Drug Test Results:

Read results at 5 minutes. A negative result will be shown by two color lines on the strip. The intensity of the line does not matter. A positive result will be shown by a single color line in the C section of the test strip. If there are no color lines present or a single color line in the T section of the test strip.

the results are invalid, and the test should be repeated.

Drug Test Results c c c c c c c c c r T T Invalid Negative Preliminary Positive Blood Alcohol Concentration

mg/dL 0.0 20 40 80 300 % 0.00% 0.00% 0.00% 0.00% 0.00%

NOTE: Results after more than 2 minutes may be not accurate



Panel Configuration Options

Item Code	Description	Complexity
Q0D0A-266-I	6-Panel T-Square with indicator; AMP, BZO, COC, MET, OPI, THC	Forensic Use Only
Q0D0A-376-A	7-Panel T-Square; AMP, COC, MET, OPI, PCP, THC + ALC	Forensic Use Only
Q0D0A-186-I	8-Panel T-Square with Indicator; AMP, BUP, BZO, COC, MET, OPI, OXY, THC	Forensic Use Only
Q0D0A-2106	10-Panel T-Square; AMP, BUP, BZO, COC, MET, MTD, OPI, OXY, PCP, THC	Forensic Use Only
Q0D0A-2106-I	10-Panel T-Square with indicator; AMP, BUP, BZO, COC, MET, MTD, OPI, OXY, PCP, THC	Forensic Use Only
Q0D0A-5106	10-Panel T -Square; AMP, BAR, BUP, BZO, COC, MET, OPI, OXY, PCP, THC	Forensic Use Only
Q0D0A-3126-A	12-Panel T -Square; AMP, BAR, BUP, BZO, COC, MET, MTD, OPI, OXY, PCP, THC+ALC	Forensic Use Only
Q0D0A-6126	12-Panel T-Square; AMP, BAR, BUP, BZO, COC, MDMA, MET, MTD, OPI, OXY, PCP, THC	Forensic Use Only
Q0D0A-1116-A	11-Panel T-Square; AMP, BAR, BUP, BZO, COC, MET, MTD, OPI, OXY, PCP + ALC	Forensic Use Only
Q0D0A-2116-A	11-Panel T-Square; AMP, BAR, BUP, BZO, COC, MET, OPI, OXY, PCP, THC+ALC	Forensic Use Only
Q0D0A-5116-I	11-Panel T-Square with Indicator; AMP, BAR, BUP, BZO, COC, MET, MDMA, MTD, OXY, OPI, THC	Forensic Use Only
Q0D0A-256EU0-I	5-Panel T-Square with indicator; AMP, COC, MET, OPI, THC	E&I Exempt
QODOA-156NTEUO-I	5-Panel T-Square with indicator; AMP, COC, MET, OPI, THC	E&I Exempt
QODOA-666EUO-I	6-Panel T-Square with indicator; AMP, COC, MET, OPI, OXY, THC	E&I Exempt
QODOA-166EUO-I	6-Panel T-Square with indicator; AMP, COC, MET, OPI, PCP, THC	E&I Exempt
QODOA-366NTEUO-I	6-Panel T-Square no THC with Indicator; AMP, COC, MET, OPI, OXY, PCP	E&I Exempt
Q0D0A-376EU0-I	7-Panel T-Square with Indicator; AMP, COC, MET, OPI, OXY, PCP, THC	E&I Exempt
QODOA-496NTEUO-I	9-Panel T-Square no THC with Indicator; AMP, BAR, COC, MDMA, MET, MTD, OPI, OXY, PCP	E&I Exempt
Q0D0A-10106EU0-I	10-Panel T-Square with indicator; AMP, BAR, COC, MDMA, MET, MTD, OPI, OXY, PCP, THC	E&I Exempt

Customizable drug panels & Alcohol test are available.